

Test Report



Model: Neck Shoulder Wrap

Temperature: 25°C

Manufacture: 660nm+850nm


Humidity: 60%

Test Date: 2023-10-25,14:11:19


Parameter

Name	Value	Name	Value	Name	Value	Name	Value
L(cd/m²)	4282.02	Centroid(nm)	767.0				
FootLamber(fl)	1249.75	Color Ratio(RGB)	100.0,0.0,0.0				
CCT(K)	1001	CIE1931 X	16284.178				
Duv	-0.01368	CIE1931 Y	6269.428				
CIE x,y	0.7220,0.2780	CIE1931 Z	0.514				
CIE u,v	0.5904,0.3410	TLCI-2012	1				
CIE u',v'	0.5904,0.5114	Integral Time(ms)	0				
SDCM	215.46	Peak Signal	56478				
Ra	-0.6	Dark Signal	2412				
Le(mW/cm²/sr)	157927.516	Compensate level	2874				
S/P	0.020	Voltage(V)	1.000				
Dominant(nm)	643.60	Current(mA)	1.000				
Purity(%)	100.0	Power(W)	1.000				
HalfWidth(nm)	19.1	Power Factor	1.000				
Peak(nm)	852.0						
Center(nm)	849.7						



Spectrogram



CIE1931



CRI



Instrument Status

Type: OHSP-350Z


Integral Time: 0.209ms

SN: 0

VPeak: 56478

Scan Range: 350-950nm

VDark: 2412



Test Report



Model: Neck Shoulder Wrap

Temperature: 25°C

Manufacture: 660nm+850nm

Humidity: 60%

Test Date: 2023-10-25,14:11:19



Power intensity

660nm:18.2、 20.6mW/cm²



850nm:28.3、 29.5mW/cm²



Instrument Status

Type: OHSP-350Z

Integral Time: 0.209ms

SN: 0

VPeak: 56478

Scan Range: 350-950nm

VDark: 2412

